

OI-922

Page 55 of 66

Rev. BJ

Vialed and Packaged Test Strip Release


Test Strips

nova
biomedical

CERTIFICATE OF COMPLIANCE Page 1/2

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	NIPRO DIAGNOSTICS INC.
Description:	TEST STRIPS, NIPRO STATSTRIP™ GLUCOSE, PACKAGED
Catalog Number:	47940
Nova Biomedical P/N	47940
Target Range (mg/dL) SD (mg/dL) /CV (%)	
30-50 ≤ 4 mg/dL	<u>PASS</u> / FAIL
51-110 ≤ 6.0%	<u>PASS</u> / FAIL
111-150 ≤ 5.0%	<u>PASS</u> / FAIL
151-250 ≤ 5.0%	<u>PASS</u> / FAIL
251-400 ≤ 5.0%	<u>PASS</u> / FAIL
Accuracy Criteria 1: Total within +/-12 mg/dL or +/- 12% of reference method ≥ 95%	<u>PASS</u> / FAIL
Accuracy Criteria 2: Total within +/-15 mg/dL or +/- 15% of reference method ≥ 98%	<u>PASS</u> / FAIL
Qty:	304
Lot No.	2123306309
Mfg. Date:	2023.11.02
Expiration Date:	2026.05.31



Quality Control Manager

11/14/23

Date

Nova Biomedical Corporation

200 Prospect Street Waltham, MA 02454

OI-922 Page 56 of 66 Rev. BJ Vial and Packaged Test Strip Release Test Strips

nova
biomedical

ADDENDUM TO CERTIFICATE OF COMPLIANCE Page 2/2

RELEASE TEST RESULTS

Customer:	NIPRO DIAGNOSTICS INC.
Description:	TEST STRIPS, NIPRO STATSTRIP™ GLUCOSE
Catalog Number:	47940
Nova Biomedical P/N	47940
Lot No.	2123306309

(Spiked whole blood) Target Range:30-50		
Mean (mg/dL) N=100	SD (mg/dL) ≤ 4	YSI (reference, mg/dL)
39	2.2	38.2
(Spiked whole blood) Target Range:51-110		
Mean (mg/dL) N=100	CV% ≤ 6.0	YSI (reference, mg/dL)
92	4.1	86.2
(Spiked whole blood) Target Range:111-150		
Mean (mg/dL) N=100	CV% ≤ 5.0	YSI (reference, mg/dL)
143	4.4	137.0
(Spiked whole blood) Target Range:151-250		
Mean (mg/dL) N=100	CV% ≤ 5.0	YSI (reference, mg/dL)
228	3.8	224.5
(Spiked whole blood) Target Range:251-400		
Mean (mg/dL) N=100	CV% ≤ 5.0	YSI (reference, mg/dL)
360	3.1	358.8

Accuracy Criteria 1: Total within +/-12 mg/dL or +/- 12% of reference method $\geq 95\%$	Accuracy Criteria 2: Total within +/-15 mg/dL or +/- 15% of reference method $\geq 98\%$
95.6%	99.6%

[Signature]
Quality Control Manager
Nova Biomedical Corporation

Date *11/14/23*
200 Prospect Street Waltham, MA 02454