



CERTIFICATE OF ANALYSIS

STP Prime ABG Cal Cartridge 300 Sample
Lot Number 23312017

Ref Number 53359
Exp Date 2025-04-26
Date of Manufacture 2023-10-26
UDI# (01)00385480533591(11)231026(17)250426(10)23312017

Cal A	L/N	23299027	
		<u>Assigned Value*</u>	<u>Result</u>
pH		7.356	N/A
pCO2 mmHg		24.3	N/A

Cal B	L/N	23299028	
		<u>Spec</u>	<u>Pass/Fail</u>
pH		6.837 - 6.843	Pass
pO2 mmHg		167.0	N/A

Flush	L/N	23299039	
		<u>Assigned Value*</u>	<u>Result</u>
pCO2 mmHg		51.5	N/A
pO2 mmHg		96.5	N/A

Reference	L/N	23299061	
		<u>Spec</u>	<u>Pass/Fail</u>
KCl		1.8 - 2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Auishi M

Approval

QC Inspector

Title

11/14/2023

Date