

CERTIFICATE OF ANALYSIS

STP CCX Cal Cartridge 4 Lot Number 23305058 Ref Number 37164 Exp Date 2025-04-11 Date of Manufacture 2023-10-11 UDI# (01)00385480371643(11)231011(17)250411(10)23305058

Cal A L/N 23285064	Cal B L/N 23284023	Cal C L/N 23284024
Spec Pass/Fail	Spec Pass/Fail	Spec Pass/Fail
pH @ 37C 7.351 - 7.357 Pass	pH @ 37C 6.822 - 6.828 Pass	Na mmol/L 139.0 - 141.0 Pass
Lac mmol/L 13.5 - 16.5 Pass	Na mmol/L 69.5 - 70.5 Pass	K mmol/L 3.97 - 4.03 Pass
	K mmol/L 9.9 - 10.1 Pass	Cl mmol/L 118.0 - 120.0 Pass
	CI mmol/L 39.7 - 40.3 Pass	BUN mg/dl 9.9 - 10.1 Pass
	Ca mmol/L 1.97 - 2.03 Pass	Glu mg/dl 78.8 - 81.2 Pass
	Mg mmol/L 1.48 - 1.52 Pass	Lac mmol/L 1.8 - 2.2 Pass
		Ca mmol/L 0.99 - 1.02 Pass
		Mg mmol/L 0.49 - 0.51 Pass
		HCO3 mmol/L 18.43 - 19.19 Pass

Cal D L/N	23284025		Flush L/N	23289060		Reference L/N	23284028	
	<u>Spec</u>	Pass/Fail		<u>Spec</u>	Result		<u>Spec</u>	Pass/Fail
BUN mg/dl	49.4 - 50.6	Pass	Aqueous Salt	N/A	Present	KCI mol/L	1.8 - 2.2	Pass
Glu mg/dl	197 - 203	Pass	Solution					
Lac mmol/L	5.5 - 6.5	Pass						
HCO3 mmol/L	37.05 - 38.57	Pass						

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

11/17/2023

Date