

Nova 4/10/11/13 Lithium Sensor Lot Number 23310019

Ref Number 15742 Exp Date 2024-11-06 Date of Manufacture 2023-11-06 UDI# (01)10385480157428(11)231106(17)241106(10)23310019

	Spec*	Pass/Fail
Slope	10.3 - 11.7	Pass
Chem Ctl L 2 mmol/L	1.77 - 2.27	Pass
Set A L 2 mmol/L	1.96 - 2.36	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

2Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector Title

11/21/2023 Date