

OI-922

Page 26 of 66

Rev. BJ

Vialed and Packaged Test Strip Release

Test Strips

nova
biomedical

CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	NIA
Description:	TEST STRIPS, STATSENSOR™ CREATININE, 50 COUNT
Catalog Number:	43370
Nova Biomedical P/N	43370
Control Level 1: CV (%) ≤ 15% Control Level 2: CV (%) ≤ 10% Control Level 3: CV (%) ≤ 10%	PASS / FAIL PASS / FAIL PASS / FAIL
Qty:	20,016
Lot No.	4923310129
Mfg. Date:	2023-11-06
Expiration Date:	2024-11-06
Unique Device Identifier (UDI):	(01)20385480433703 (11)231106 (17)241106 (10)4923310129

Marc Broome
Quality Control Manager

11/21/2023
Date

Nova Biomedical Corporation

200 Prospect Street

Waltham, MA 02454