

CERTIFICATE OF ANALYSIS

Allegro HbA1c Test Cartridge Lot Number 23321029

tridge Ref Number 54215 Exp Date 2024-11-17 Date of Manufacture 2023-11-17 UDI# (01)00385480542159(11)231117(17)241117(10)23321029

	Ctl Level 1			Ctl Level 2		
	<u>Spec</u>	Pass/Fail		<u>Spec</u>	Pass/Fail	
HbA1c %	4.8 - 6.6	Pass	HbA1c %	7.9 - 11.5	Pass	

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Frit

Approval

<u>QC Inspector</u> Title

<u>11/21/2023</u> Date