



CERTIFICATE OF ANALYSIS

Allegro Lipids Test Cartridge
Lot Number 23320001

Ref Number 54652
Exp Date 2025-05-16
Date of Manufacture 2023-11-16
UDI# (01)00385480546522(11)231116(17)250516(10)23320001

Ctl Level 1			Ctl Level 2		
	<u>Spec</u>	<u>Pass/Fail</u>		<u>Spec</u>	<u>Pass/Fail</u>
HDL mg/dL	34 - 56	Pass	HDL mg/dL	68 - 102	Pass
TCHOL mg/dL	156 - 226	Pass	TCHOL mg/dL	285 - 401	Pass
TGL mg/dL	58 - 134	Pass	TGL mg/dL	250 - 390	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Haneen Aldarwish

Approval

QC Inspector
Title

11/22/2023
Date