

CERTIFICATE OF ANALYSIS

Nova StatStrip Glucose Control Solution Level 2

Nova Reference 54521 LifeScan Reference 02315502 Lot Number 0423284242 Date of Manufacture 2023-10-11 Exp Date 2025-10-11 UDI# (01)00385480545211(11)231011(17)251011(10)0423284242

Mean

<u>Spec</u>

Result

Glucose mg/dL

105 - 125

116

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Amishi.M

Title

11/27/2023

Date