



**CERTIFICATE OF ANALYSIS**  
**StatStrip Glucose Control Solution Level 3**

Nova Ref Number 54522  
LifeScan Ref Number 02315602  
Lot Number 0423285243

Date of Manufacture 2023-10-12  
Exp Date 2025-10-12  
UDI # (01)00385480545228(11)231012(17)251012(10)0423285243

|               | <u>Spec</u> | <u>Mean Result</u> |
|---------------|-------------|--------------------|
| Glucose mg/dL | 275 - 325   | 306                |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

*Haneen Aldarwish*

Approval

QC Inspector  
Title

11/21/2023  
Date