



CERTIFICATE OF ANALYSIS

STP CCX Cal Cartridge 3
Lot Number 23297085

Ref Number 36640
Exp Date 2025-04-02
Date of Manufacture 2023-10-02
UDI# (01)00385480366403(11)231002(17)250402(10)23297085

Cal A L/N 23276090	Cal B L/N 23275037	Cal C L/N 23275038
<u>Spec</u> <u>Pass/Fail</u>	<u>Spec</u> <u>Pass/Fail</u>	<u>Spec</u> <u>Pass/Fail</u>
pH @ 37C 7.351 - 7.357 Pass	pH @ 37C 6.822 - 6.828 Pass	Na mmol/L 139.0 - 141.0 Pass
Lac mmol/L 13.5 - 16.5 Pass	Na mmol/L 69.5 - 70.5 Pass	K mmol/L 3.97 - 4.03 Pass
	K mmol/L 9.9 - 10.1 Pass	Cl mmol/L 118.0 - 120.0 Pass
	Cl mmol/L 39.7 - 40.3 Pass	BUN mg/dl 9.9 - 10.1 Pass
	Ca mmol/L 1.97 - 2.03 Pass	Glu mg/dl 78.8 - 81.2 Pass
	Mg mmol/L 1.48 - 1.52 Pass	Lac mmol/L 1.8 - 2.2 Pass
		Ca mmol/L 0.99 - 1.02 Pass
		Mg mmol/L 0.49 - 0.51 Pass
		HCO3 mmol/L 18.43 - 19.19 Pass

Cal D L/N 23275039	Flush L/N 23277041	Reference L/N 23275027
<u>Spec</u> <u>Pass/Fail</u>	<u>Spec</u> <u>Result</u>	<u>Spec</u> <u>Pass/Fail</u>
BUN mg/dl 49.4 - 50.6 Pass	Aqueous Salt N/A Present	KCl mol/L 1.8 - 2.2 Pass
Glu mg/dl 197 - 203 Pass	Solution	
Lac mmol/L 5.5 - 6.5 Pass		
HCO3 mmol/L 37.05 - 38.57 Pass		

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Haneen Aldarwish

Approval

QC Inspector
Title

11/29/2023
Date