

CERTIFICATE OF ANALYSIS

Bio pHOx Cal Pack Lot Number 23332047 Part Number 41874 Exp Date 2025-05

Std A L/N 23321045

Std B L/N 23321046

Std C L/N 23317054

 Spec
 Pass/Fail

 Na mmol/L 139.0 - 141.0
 Pass

 HCO3 mmol/L 19.6 - 20.4
 Pass

Spec Pass/Fail HCO3 mmol/L 39.2 - 40.8 Pass

<u>Spec</u> <u>Pass/Fail</u> pH @ 37C 7.352 - 7.358 Pass

Std D L/N 23319011

Reference L/N 23317026

<u>Spec Pass/Fail</u> Cl mol/l 1 8 - 2 2 Pass

302 - 6.808 Pass KCI mol/L 1.8 - 2.2 Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: Na, HCO3, pH, KCI.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Amishi.M

Approval

QC Inspector Title

11/30/2023 Date