

CERTIFICATE OF ANALYSIS

 STAT PROFILE® pHOx Ultra™/CCX™ Potassium Sensor
 Ref Number 48582

 Lot Number 23318087
 Exp Date 2025-02-14

 Date of Manufacture 2023-11-14
 UDI# (01)00385480485821(11)231114(17)250214(10)23318087

	<u>Spec</u>	Pass/Fail
Slope	9.00 to 11.00	Pass
Std C mV	0 + 50.00	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Tymolly

Approval

<u>QC Inspector</u> Title

<u>12/1/2023</u> Date