

CERTIFICATE OF ANALYSIS

Bioprofile 300 Lactate Membrane Lot Number 23339004

Part Number 34573 Exp Date 2024-12

| | <u>Spec*</u> | Mean Result |
|----------------|--------------|-------------|
| Slope | 10.0 - 60.0 | 40.67 |
| Biopro Ctl L 6 | .75 - 1.25 | 0.87 |
| Biopro Ctl L 7 | 2.50 - 3.50 | 2.79 |

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

andang

Approval

QC Inspector Title

<u>12/8/2023</u> Date