



CERTIFICATE OF ANALYSIS

BioPro Flex System Fluid Cartridge
Lot Number 23340062

Part Number 39580
Exp Date 2025-01

	S Sol L/N	23286050	
	<u>Spec</u>		<u>Pass/Fail</u>
pH	6.900 - 7.100		Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: pH.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

12/8/2023

Date