



CERTIFICATE OF ANALYSIS

Nova 1/4/5/8/10/11/12/13/14 Sodium Sensor
Lot Number 23334061

Ref Number 14714
Exp Date 2025-05-30
Date of Manufacture 2023-11-30
UDI# (01)00385480147149(11)231130(17)250530(10)23334061

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	9.8 - 10.9	Pass
Std A mV	-10.00 - +70.00	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Approval

QC Inspector

Title

12/12/2023

Date