



CERTIFICATE OF ANALYSIS

Nova 1-11, 13 Dilutor Pack
Lot Number 23314006

Ref Number 12882
Exp Date 2024-11-10
Date of Formulation 2023-11-10
UDI# (01)00385480128827(11)231110(17)241110(10)23314006

Cal L L/N 23314006

		<u>Spec</u>	<u>Pass/Fail</u>
Na	mmol/L	29.8 - 30.2	Pass
K	mmol/L	9.9 - 10.1	Pass
Cl	mmol/L	39.7 - 40.3	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Approval

QC Inspector

Title

12/13/2023

Date