



CERTIFICATE OF ANALYSIS

Nova 5 Calibrator Pack
Lot Number 330712

Ref Number 15648
Exp Date 2025-11-03
Date of Manufacture 2023-11-03
UDI# (01)00385480156486(11)231103(17)251103(10)330712

Cal A L/N	23307012	Cal B L/N	23307012	Reference L/N	23307012			
	<u>Spec</u>	<u>Pass/fail</u>	<u>Spec</u>	<u>Pass/fail</u>	<u>Spec</u>	<u>Pass/fail</u>		
Na mmol/L	139.0 - 141.0	Pass	Na mmol/L	59.4 - 60.6	Pass	KCl mol/L	1.8 - 2.2	Pass
K mmol/L	3.97 - 4.03	Pass	K mmol/L	9.8 - 10.2	Pass			
Cl mmol/L	113.6 - 115.4	Pass	Cl mmol/L	45.5 - 46.5	Pass			

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Haneen Aldarwish

Approval

QC Inspector
Title

12/13/2023
Date