



CERTIFICATE OF ANALYSIS

STP Prime ABG Cal Cartridge 400 Sample
Lot Number 23340029

Ref Number 53463
Exp Date 2025-05-22
Date of Manufacture 2023-11-22
UDI# (01)00385480534635(11)231122(17)250522(10)23340029

| | | | | |
|-----------|-----|------------------------|---------------|--|
| Cal A | L/N | 23326017 | | |
| | | <u>Assigned Value*</u> | <u>Result</u> | |
| pH | | 7.359 | N/A | |
| pCO2 mmHg | | 24.4 | N/A | |

| | | | | |
|----------|-----|---------------|------------------|--|
| Cal B | L/N | 23326018 | | |
| | | <u>Spec</u> | <u>Pass/Fail</u> | |
| pH | | 6.837 - 6.843 | Pass | |
| pO2 mmHg | | 167.0 | N/A | |

| | | | | |
|-----------|-----|------------------------|---------------|--|
| Flush | L/N | 23326015 | | |
| | | <u>Assigned Value*</u> | <u>Result</u> | |
| pCO2 mmHg | | 50.7 | N/A | |
| pO2 mmHg | | 96.8 | N/A | |

| | | | | |
|-----------|-----|-------------|------------------|--|
| Reference | L/N | 23326014 | | |
| | | <u>Spec</u> | <u>Pass/Fail</u> | |
| KCl | | 1.8 - 2.2 | Pass | |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Haneen Aldarwish

Approval

QC Inspector

Title

12/18/2023

Date