



CERTIFICATE OF ANALYSIS

StatStrip Glucose Control Level 3

LifeScan Reference 02315601

Lot Number 0422154303

Date of Manufacture 2022-06-03

Exp Date 2024-12-03

	<u>Spec</u>	<u>Mean Result</u>
Glucose mg/dL	275-325	309

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

7/6/2022

Date