



CERTIFICATE OF ANALYSIS

Nova Glu/Lac Sensor Soaking Solution
Lot Number 23346032

Ref Number 13409

Exp Date 2024-12-12

Date of Manufacture 2023-12-12

UDI# (01)00385480134095(11)231212(17)241212(10)23346032

	<u>Spec</u>	<u>Pass/Fail</u>
NaOCl	0.43 - 0.73%	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Approval

QC Inspector

Title

12/20/2023

Date