



CERTIFICATE OF ANALYSIS

Allegro Lipids Test Cartridge
Lot Number 23340018

Ref Number 54652
Exp Date 2025-06-06
Date of Manufacture 2023-12-06
UDI# (01)00385480546522(11)231206(17)250606(10)23340018

| Ctl Level 1 | | | Ctl Level 2 | | |
|-------------|-------------|------------------|-------------|-------------|------------------|
| | <u>Spec</u> | <u>Pass/Fail</u> | | <u>Spec</u> | <u>Pass/Fail</u> |
| HDL mg/dL | 34 - 56 | Pass | HDL mg/dL | 68 - 102 | Pass |
| TCHOL mg/dL | 156 - 226 | Pass | TCHOL mg/dL | 285 - 401 | Pass |
| TGL mg/dL | 58 - 134 | Pass | TGL mg/dL | 250 - 390 | Pass |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Haneen Aldarwish

Approval

QC Inspector
Title

12/27/2023
Date