

CERTIFICATE OF ANALYSIS

Allegro Lipids Test Cartridge Lot Number 23345015 Ref Number 54652 Exp Date 2025-06-11 Date of Manufacture 2023-12-11 UDI# (01)00385480546522(11)231211(17)250611(10)23345015

Ctl Level 1

Ctl Level 2

	<u>Spec</u>	Pass/Fail		<u>Spec</u>	Pass/Fail
HDL mg/dL TCHOL mg/dL	34 - 56 156 - 226	Pass Pass	HDL mg/dL TCHOL mg/dL	68 - 102 285 - 401	Pass Pass
TGL mg/dL	58 - 134	Pass	TGL mg/dL	250 - 390	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

fuman Achta

Approval

QC Inspector Title

<u>12/26/2023</u> Date