

OI-922

Page 47 of 65

Rev. BG

Vialed and Packaged Test Strip Release

Test Strips

*nova*  
biomedical

**CERTIFICATE OF COMPLIANCE**

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	N/A	
Description:	TEST STRIPS, STATSTRIP HEMOGLOBIN/HEMATOCRIT, PACKAGED 50 CT	
Catalog Number:	58786	
Nova Biomedical P/N	58786	
<u>Target Range (Hct)</u>	<u>CV (%)</u>	
28-36 %	≤ 8%	PASS / FAIL
38-46 %	≤ 8%	PASS / FAIL
48-56%	≤ 8%	PASS / FAIL
Qty:	171	
Lot No.	1522185249	
Date of Mfg. (DOM):	2022 / 07 / 05 (YYYY/MM/DD)	
Expiration Date:	2024 / 07 / 04 (YYYY/MM/DD)	
Unique Device Identifier (UDI):	(01)20385480587864 (11)220705 (17)240704 (10)1522185249	

Mark Brosme  
Quality Control Manager

7/13/2022  
Date

Nova Biomedical Corporation

200 Prospect Street Waltham, MA 02454