



CERTIFICATE OF ANALYSIS

STP® pHox Ultra™ / CCX™ Chloride Sensor
Lot Number 23345021

Part Number 48557
Exp Date 2025-03-11

Date of Manufacture 2023-12-11

UDI# (01)10385480485576(11)231211(17)250311(10)23345021

| | <u>Spec</u> | <u>Pass/Fail</u> |
|-------------|---------------|------------------|
| Slope | 8.50 - 13.00 | Pass |
| pHOx Ctl L4 | 121.3 - 130.3 | Pass |
| pHOx Ctl L5 | 91.5 - 100.5 | Pass |
| Std C mV | 10.00 - 90.00 | Pass |

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

1/2/2024
Date