



CERTIFICATE OF ANALYSIS

STP Prime ABG Cal Cartridge 300 Sample  
Lot Number 23355035

Ref Number 53359  
Exp Date 2025-05-30  
Date of Manufacture 2023-11-30  
UDI# (01)00385480533591(11)231130(17)250530(10)23355035

Cal A	L/N	23334012	
		<u>Assigned Value*</u>	<u>Result</u>
	pH	7.352	N/A
	pCO2 mmHg	24.9	N/A

Cal B	L/N	23334013	
		<u>Spec</u>	<u>Pass/Fail</u>
	pH	6.837 - 6.843	Pass
	pO2 mmHg	167.0	N/A

Flush	L/N	23334010	
		<u>Assigned Value*</u>	<u>Result</u>
	pCO2 mmHg	52.2	N/A
	pO2 mmHg	97.2	N/A

Reference	L/N	23335038	
		<u>Spec</u>	<u>Pass/Fail</u>
	KCl	1.8 - 2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

01/02/2024  
Date