

OI-922

Page 48 of 66

Rev. BH

Vialed and Packaged Test Strip Release

Test Strips

nova
biomedical

CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	N/A												
Description:	TEST STRIPS, STATSTRIP HEMOGLOBIN/HEMATOCRIT, PACKAGED 50 CT												
Catalog Number:	58786												
Nova Biomedical P/N	58786												
<table border="1"><thead><tr><th>Target Range (Hct)</th><th>CV (%)</th></tr></thead><tbody><tr><td>28-36 %</td><td>≤ 8%</td></tr><tr><td>38-46 %</td><td>≤ 8%</td></tr><tr><td>48-56%</td><td>≤ 8%</td></tr></tbody></table>	Target Range (Hct)	CV (%)	28-36 %	≤ 8%	38-46 %	≤ 8%	48-56%	≤ 8%	<table border="1"><tbody><tr><td>PASS / FAIL</td></tr><tr><td>PASS / FAIL</td></tr><tr><td>PASS / FAIL</td></tr></tbody></table>		PASS / FAIL	PASS / FAIL	PASS / FAIL
Target Range (Hct)	CV (%)												
28-36 %	≤ 8%												
38-46 %	≤ 8%												
48-56%	≤ 8%												
PASS / FAIL													
PASS / FAIL													
PASS / FAIL													
Qty:	1,458												
Lot No.	1523044249												
Date of Mfg. (DOM):	2023 / 02 / 13 (YYYY/MM/DD)												
Expiration Date:	2025 / 02 / 13 (YYYY/MM/DD)												
Unique Device Identifier (UDI):	(01)20385480587864 (11)230213 (17)250213 (10)1523044249												

Quality Control Manager

Date

3/1/23

Nova Biomedical Corporation

200 Prospect Street Waltham, MA 02454