

OI-922 Page 55 of 66 Rev. BJ Vial and Packaged Test Strip Release Test Strips

nova
biomedical

CERTIFICATE OF COMPLIANCE Page 1/2

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	NIPRO DIAGNOSTICS INC.	
Description:	TEST STRIPS, NIPRO STATSTRIP™ GLUCOSE, PACKAGED	
Catalog Number:	47940	
Nova Biomedical P/N	47940	
Target Range (mg/dL)	SD (mg/dL) /CV (%)	
30-50	≤ 4 mg/dL	PASS / FAIL
51-110	≤ 6.0%	PASS / FAIL
111-150	≤ 5.0%	PASS / FAIL
151-250	≤ 5.0%	PASS / FAIL
251-400	≤ 5.0%	PASS / FAIL
Accuracy Criteria 1: Total within +/-12 mg/dL or +/- 12% of reference method ≥ 95%		PASS / FAIL
Accuracy Criteria 2: Total within +/-15 mg/dL or +/- 15% of reference method ≥ 98%		PASS / FAIL
Qty:	304	
Lot No.	2123346309	
Mfg. Date:	2023.12.12	
Expiration Date:	2026.06.30	


Quality Control Manager

1/4/24
Date

Nova Biomedical Corporation 200 Prospect Street Waltham, MA 02454

OI-922 Page 56 of 66 Rev. BJ Viald and Packaged Test Strip Release Test Strips

nova
biomedical


ADDENDUM TO CERTIFICATE OF COMPLIANCE Page 2/2


RELEASE TEST RESULTS

Customer:	NIPRO DIAGNOSTICS INC.
Description:	TEST STRIPS, NIPRO STATSTRIP™ GLUCOSE
Catalog Number:	47940
Nova Biomedical P/N	47940
Lot No.	2123346309

(Spiked whole blood) Target Range:30-50		
Mean (mg/dL) N=100	SD (mg/dL) ≤ 4	YSI (reference, mg/dL)
44	2.1	45.6
(Spiked whole blood) Target Range:51-110		
Mean (mg/dL) N=100	CV% ≤ 6.0	YSI (reference, mg/dL)
97	2.7	95.9
(Spiked whole blood) Target Range:111-150		
Mean (mg/dL) N=100	CV% ≤ 5.0	YSI (reference, mg/dL)
139	3.1	139.0
(Spiked whole blood) Target Range:151-250		
Mean (mg/dL) N=100	CV% ≤ 5.0	YSI (reference, mg/dL)
228	2.6	227.3
(Spiked whole blood) Target Range:251-400		
Mean (mg/dL) N=100	CV% ≤ 5.0	YSI (reference, mg/dL)
370	3.2	374.3

Accuracy Criteria 1: Total within +/-12 mg/dL or +/- 12% of reference method $\geq 95\%$	Accuracy Criteria 2: Total within +/-15 mg/dL or +/- 15% of reference method $\geq 98\%$
100.0 %	100.0 %


Quality Control Manager
Nova Biomedical Corporation


Date
200 Prospect Street Waltham, MA 02454