

CERTIFICATE OF ANALYSIS

STAT PROFILE® pHOx Ultra™/CCX™ pO2 Sensor Lot Number 23348046 Ref Number 48598 Exp Date 2025-12-14

Date of Manufacture 2023-12-14 UDI# (01)00385480485982(11)231214(17)251214(10)23348046

<u>Spec</u>

Pass/Fail

Slope

-3.00 to -6.00

Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

1/9/2024

Date