



CERTIFICATE OF ANALYSIS

StatStrip Glucose Control Level 3

LifeScan Reference 02315601

Lot Number 0422236303

Date of Manufacture 2022-08-24

Exp Date 2025-02-24

	<u>Spec</u>	<u>Mean Result</u>
Glucose mg/dL	275-325	311

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Awishi M

Approval

QC Inspector

Title

9/20/2022

Date