



**CERTIFICATE OF ANALYSIS**

STAT PROFILE® pHOx Ultra™/CCX™ Potassium Sensor  
Lot Number 23362035

Ref Number 48582

Exp Date 2025-03-28

Date of Manufacture 2023-12-28

UDI# (01)00385480485821(11)231228(17)250328(10)23362035

|          | <u>Spec</u>   | <u>Pass/Fail</u> |
|----------|---------------|------------------|
| Slope    | 9.00 to 11.00 | Pass             |
| Std C mV | 0 ± 50.00     | Pass             |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

1/17/2024

Date