



**CERTIFICATE OF ANALYSIS**

STP® pHox Ultra™ / CCX™ Calcium Sensor  
Lot Number 23361117

Part Number 48581  
Exp Date 2025-03-27  
Date of Manufacture 2023-12-27  
UDI# (01)00385480485814(11)231227(17)250327(10)23361117

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	8.50 - 12.50	Pass
Std C mV	0.00 ± 50.00	Pass
pHox Ctl L4	1.03 - 1.19	Pass
pHox Ctl L5	1.51 - 1.75	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

1/17/2024

Date