



**CERTIFICATE OF ANALYSIS**

StatSensor Creatinine Control Solution Level 2  
Lot Number 5023360242

Ref Number 43922  
Exp Date 2025-12-26  
Date of Manufacture 2023-12-26  
UDI# (01)00385480439220(11)231226(17)251226(10)5023360242

	<u>Spec</u>	<u>Mean Result</u>
Creatinine mg/dL	1.95 - 2.25	2.07

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Haneen Aldarwish

Approval

QC Inspector

Title

1/25/2024

Date