



CERTIFICATE OF ANALYSIS

Bioprofile Control Level 7
Lot Number 336121

Catalog Number 33837
Exp Date 2025-06

	Spec	Pass/Fail
pH	6.97 - 7.03	Pass
Gluc g/L	9.40-10.60	Pass
Ace mmol/L	80.0-120.0	Pass
PO4 mmol/L	3.60-4.40	Pass
Gly g/L	4.50-5.50	Pass
NH4+ mmol/L	9.4-10.6	Pass
Na+ mmol/L	90-100	Pass
K+ mmol/L	9.2-10.8	Pass
Lac g/L	2.70-3.30	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Awishi M

Approval

QC Inspector
Title

1/26/2024
Date