



CERTIFICATE OF ANALYSIS

PO2 Internal Filling Solution
Lot Number 24016085

Ref Number 06554
Exp Date 2025-07-16
Date of Manufacture 2024-01-16

	<u>Spec</u>	<u>Pass/Fail</u>
Na mM	60.0 - 80.0	Pass
K mM	27.0 - 35.0	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Auishi M

Approval

QC Inspector
Title

1/26/2024
Date