



CERTIFICATE OF ANALYSIS

Bioprofile Control Level 6
Lot Number 336120

Catalog Number 33836
Exp Date 2025-06

	Spec	Pass/Fail
pH	7.55 - 7.61	Pass
Gluc g/L	1.80-2.20	Pass
Ace mmol/L	17.0-23.0	Pass
PO4 mmol/L	0.85-1.15	Pass
Gly g/L	2.20-2.80	Pass
NH4+ mmol/L	1.7-2.3	Pass
Na+ mmol/L	135-145	Pass
K+ mmol/L	3.6-4.4	Pass
Lac g/L	0.90-1.10	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Awishi M

Approval

QC Inspector
Title

2/5/2024
Date