



CERTIFICATE OF ANALYSIS

BioProfile CDV Reagent Pack
Lot Number 24015026

Ref Number 48796
Exp Date 2024-12

CDV Bottle Pack L/N 24015025			Exp Date 2024-12		
C Soln L/N 23349094			T Soln L/N 23350005		
	<u>Spec</u>	<u>Result</u>		<u>Spec</u>	<u>Result</u>
3% Detergent	N/A	Present	0.4% Trypan	N/A	Present
pH	> 9.0	Pass	Particle Count	≤ 2.0	Pass
CDV Cartridge L/N 24015024			Exp Date 2024-12		
S Soln L/N 23355042			D Soln L/N 23350003		
	<u>Spec</u>	<u>Pass/Fail</u>		<u>Spec</u>	<u>Result</u>
pH	6.900 - 7.100	Pass	%NaOCl	0.43 - 0.73	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: pH.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Auishi M

Approval

QC Inspector

Title

2/6/2024

Date