



CERTIFICATE OF ANALYSIS

Allegro Control 1 UACR
Lot Number 8524010241
Ref Number 51032
Exp Date 2026-01-10
Date of Manufacture 2024-01-10
UDI# (01)00385480510325(11)240110(17)260110(10)8524010241

	Spec	Pass/Fail
Albumin, mg/dL	33-81	Pass
Creatinine, mg/dL	65-119	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

2/14/2024
Date