



CERTIFICATE OF ANALYSIS

STP pHox Ultra/CCX Lactate Membrane
Lot Number 24040003

Part Number 48554
Exp Date 2025-05-09
Date of Manufacture 2024-02-09
UDI# (01)00385480485548(11)240209(17)250509(10)24040003

	<u>Spec*</u>	<u>Mean Result</u>
Slope 2	≥ 15.00	50.80
pHOx Ultra/CCX Ctl L4	1.7 - 2.3	2.0
pHOx Ultra/CCX Ctl L5	6.3 - 7.7	7.3
Lin Soln Set A Level 1	14.0 - 20.0	17.9

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

2/22/2024
Date