










5

Assay Data Sheet - Expected Ranges
BioProfile® FLEX Controls - LEVEL 5

5

Lot Number:  402732		Exp. Date: 2025-07	
Parameter	Units	Expected Range	
		Lower Limit	Upper Limit
Gln	mmol/L	 4.37	 5.87
Glu	mmol/L	 1.78	 2.38
Gluc	g/L	 8.05	 12.05
Lac	g/L	 4.42	 5.92



200 Prospect Street • Waltham, MA U.S.A. 02454-9141 • (781) 864-0800