



CERTIFICATE OF ANALYSIS

STP Prime ABG Cal Cartridge 100 Sample
Lot Number 24046062

Ref Number 53104
Exp Date 2025-08-01
Date of Manufacture 2024-02-01
UDI# (01)00385480531047(11)240201(17)250801(10)24046062

Cal A	L/N	24032005	
		<u>Assigned Value*</u>	<u>Result</u>
pH		7.359	N/A
pCO2 mmHg		24.4	N/A

Cal B	L/N	24032006	
		<u>Spec</u>	<u>Pass/Fail</u>
pH		6.837 - 6.843	Pass
pO2 mmHg		167.0	N/A

Flush	L/N	24032003	
		<u>Assigned Value*</u>	<u>Result</u>
pCO2 mmHg		52.2	N/A
pO2 mmHg		95.7	N/A

Reference	L/N	24032007	
		<u>Spec</u>	<u>Pass/Fail</u>
KCl		1.8 - 2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Haneen Aldarwish

Approval

QC Inspector
Title

02/27/2024
Date