

CERTIFICATE OF ANALYSIS

Prime Cal Cartridge ES Comp 400 Sample Lot Number 24052091 Ref Number 53460 Exp Date 2025-08-14

Date of Manufacture 2024-02-14

UDI# (01)00385480534604(11)240214(17)250814(10)24052091

Cal A L/N 24045032		Cal E	Cal B L/N 24045033			Reference L/N 24045017		
<u>Spec</u>	Pass/Fail		Spec	<u>C</u>	Pass/Fail		<u>Spec</u>	Pass/Fail
Na mmol/L 134.1 - 135.9	Pass	Na mmol/L	76.5 - 7	7.5	Pass	KCI mol/L	1.8 - 2.2	Pass
K mmol/L 4.02 - 4.08	Pass	K mmol/L	9.80 - 1	0.20	Pass			
CI mmol/L 114.1 - 115.9	Pass	CI mmol/L	44.6 - 4	15.5	Pass			
iCa mmol/L 0.99 - 1.01	Pass	iCa mmol/L	1.98 - 2	2.02	Pass			
iMg mmol/L 0.40 - 0.50	Pass	iMg mmol/L	1.48 - 1	.52	Pass			

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Haneen Aldarwish

Approval

QC Inspector Title

02/27/2024 Date