OI-922 Page 39 of 66 Rev. BJ Vialed and Packaged Test Strip Release

Test Strips



## **CERTIFICATE OF COMPLIANCE**

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	LIFESCAN
Description:	TEST STRIPS, STATSTRIP <sup>™</sup> GLUCOSE, PACKAGED 18 BOXES OF 100 COUNT
Lifescan Purchase Order No.:	637902
Catalog Number:	42214
Nova Biomedical P/N	54515
One Touch P/N	02315302
Target Range (mg/dL) SD(mg/dL) /CV (%)	7
30-50 ≤4 mg/dL	PASS FAIL
51-110 ≤ 6.0%	PASS FAIL
111-150 ≤ 5.0%	PASS/ FAIL
151-250 ≤ 5.0%	PASS FAIL
251-400 ≤ 5.0%	(PASS) FAIL
Accuracy Criteria 1: Total within +/-12 mg/dL or +/- 12% of reference method ≥ 95%	PASS/ FAIL
Accuracy Criteria 2: Total within +/-15 mg/dL or +/- 15% of reference method $\geq$ 98%	PASS) FAIL
Qty:	21,108
Lot No.	0324051249
Date of Mfg. (DOM):	2024/02/20 (YYYY/MM/DD)
Expiration Date:	2024/02/20 (YYYY/MM/DD)
Unique Device Identifier (UDI):	(01)20385480422141
	$(11)$ $\frac{2}{3}$ $\frac{4}{3}$ $\frac{0}{3}$ $\frac{2}{3}$ $\frac{0}{3}$
	(17) <u>2 6 0 2 2 0</u> (10) 0 3 2 4 0 5 1 2 4 9
In A	

Nova Biomedical Corporation 200 Prospect St

Waltham, MA 02454