

STP® pHOx Ultra™ / CCX™ BUN Sensor Lot Number 24039047

r Part Number 48555 Exp Date 2025-05-08 Date of Manufacture 2024-02-08 UDI# (01)00385480485555(11)240208(17)250508(10)24039047

	<u>Spec</u>	Pass/Fail
Slope	<u>></u> 9.60	Pass
pHOx Ctl L4	14 - 24	Pass
pHOx Ctl L5	49 - 63	Pass
Std D mV	0.00 <u>+</u> 60.00	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

3/4/2024 Date