



CERTIFICATE OF ANALYSIS

Stat Profile pHox Plus L Calibrator Cartridge C
Lot Number 24060042

Ref Number 34087
Exp Date 2025-08-23

Date of Manufacture 2024-02-23

UDI# (01)00385480340878(11)240223(17)250823(10)24060042

Table with 3 columns: Analyte, Spec, Pass/Fail. Rows include Cal-A L/N 24054002, Cal-B L/N 24054003, and Cal-C L/N 24054004 with various analytes like Na, HCO3, K, Cl, Ca, Glu, Lac.

Table for Cal-D L/N 24054005 with columns Spec and Pass/Fail. Analytes include pH @ 37C, Glu mg/dL, and Lac mmol/L.

Table for Reference L/N 24054001 with columns Spec and Pass/Fail. Analyte is KCl mol/L.

This certifies that this product was manufactured and tested at Nova Biomedica-Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical-Devices Quality Management Systems Requirements, Medical-Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Handwritten signature of Luciano A. S. A.

Approval

QC Inspector
Title

3/4/2024
Date