



**CERTIFICATE OF ANALYSIS**

Nova 1/4/5/8/10/11/12/13/14 Reference Sensor  
Lot Number 23209049

Part Number 14244  
Exp Date 2025-01-28

Date of Manufacture 2023-07-28

UDI# (01)00385480142441(11)230728(17)250128(10)23209049

	<u>Spec*</u>	<u>Pass/Fail</u>
Na Slope	$\geq 9.8$	Pass
Set A Level 3	137.2 - 143.2	Pass
CI Slope	$\geq 10.5$	Pass
Set A Level 3	110.4 - 118.4	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

<sup>2</sup>Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

3/5/2024

Date