

OI-922

Page 65 of 66

Rev. BJ

Vialed and Packaged Test Strip Release

Test Strips

nova
biomedical

CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	NIA															
Description:	TEST STRIPS, STATSTRIP LACTATE, PACKAGED															
Catalog Number:	47486															
Nova Biomedical P/N	47486															
<table border="1"><thead><tr><th>Target Range</th><th>CV (%)</th><th>SD (mmol/L)</th></tr></thead><tbody><tr><td>0.8-2.5 mmol/L</td><td>≤ 6.5%</td><td>±0.3</td></tr><tr><td>3.0-5.0 mmol/L</td><td>≤ 5.5%</td><td>±0.5</td></tr><tr><td>10-14 mmol/L</td><td>≤ 5.0%</td><td>±1.2</td></tr></tbody></table>	Target Range	CV (%)	SD (mmol/L)	0.8-2.5 mmol/L	≤ 6.5%	±0.3	3.0-5.0 mmol/L	≤ 5.5%	±0.5	10-14 mmol/L	≤ 5.0%	±1.2	<table border="1"><tbody><tr><td>PASS / FAIL</td></tr><tr><td>PASS / FAIL</td></tr><tr><td>PASS / FAIL</td></tr></tbody></table>	PASS / FAIL	PASS / FAIL	PASS / FAIL
Target Range	CV (%)	SD (mmol/L)														
0.8-2.5 mmol/L	≤ 6.5%	±0.3														
3.0-5.0 mmol/L	≤ 5.5%	±0.5														
10-14 mmol/L	≤ 5.0%	±1.2														
PASS / FAIL																
PASS / FAIL																
PASS / FAIL																
Qty:	7.398															
Lot No.	2624051249															
Mfg. Date:	2024 - 02 - 20															
Expiration Date:	2026 - 02 - 20															
Unique Device Identifier (UDI):	(01)20385480 474867 (11) 240220 (17) 260220 (10) 2624051249															

Marc Broome
Quality Control Manager

3/8/2024
Date

Nova Biomedical Corporation

200 Prospect Street

Waltham, MA 02454