



CERTIFICATE OF ANALYSIS

NCRT 4 Calibrator Pack  
Lot Number 407144

Ref Number 15647  
Exp Date 2025-05-22  
Date of Manufacture 2024-02-22  
UDI#(01)00385480156479(11)240222(17)250522(10)407144

CAI A L/N 24061002		
	<u>Spec</u>	<u>Pass/Fail</u>
Na mmol/L	129.1-130.9	Pass
K mmol/L	3.97-4.03	Pass
Cl mmol/L	115.1-116.9	Pass
HCO3 mmol/L	19.2-20.8	Pass
Li mmol/L	0.49-0.51	Pass

Cal B L/N 24061003		
	<u>Spec</u>	<u>Pass/Fail</u>
Na mmol/L	59.4-60.6	Pass
K mmol/L	9.8-10.2	Pass
Cl mmol/L	60.4-61.6	Pass
HCO3 mmol/L	9.6-10.4	Pass
Li mmol/L	4.95-5.05	Pass

Cal H L/N 24060032		
	<u>Spec</u>	<u>Result</u>
CO2 Releasing Agent	N/A	Present

Reference L/N 24053048		
	<u>Spec</u>	<u>Pass/Fail</u>
KCl mol/L	1.8-2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Approval

QC Inspector  
Title

3/12/2024  
Date