



**CERTIFICATE OF ANALYSIS**

Nova StatStrip Lactate Control Solution Level 1  
Lot Number 2724057301

Ref Number 47553  
Exp Date 2026-08-26  
Date of Manufacture 2024-02-06  
UDI# (01)00385480475532(11)240226(17)260826(10)2724057301

	<u>Spec</u>	<u>Mean Result</u>
Lactate mmol/L	0.5-0.7	0.6

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

3/12/2024  
Date