

CERTIFICATE OF ANALYSIS

PCO₂ / TCO₂ CONDITIONING SOLUTION Lot Number 24004058

LUTION Part Number 06857

Exp Date 2025-04-04

Date of Manufacturing 2024-01-04

UDI# (01)00385480068574(11)240104(17)250404(10)24004058

<u>Spec</u> <u>Pass/Fail</u> pH 0.70 - 1.30 Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: pH.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Approval

QC Inspector

3/13/2024 Date