



CERTIFICATE OF ANALYSIS

Bioprofile Flex Na Sensor  
Lot Number 24052026

Part Number 42462  
Exp Date 2025-08

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	9.0 - 10.5	Pass
Std C mV	-80.00 - +80.00	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to test specifications.

Approval

QC Inspector  
Title

3/12/2024  
Date